

**BIOPHEN LMWH CONTROL PLASMA**

Human plasmas at two levels of Low Molecular Weight Heparin (LMWH) for the quality control of Heparin measurements with anti-Xa methods

**REF** 223001

For in vitro diagnostic use only

**LOT** F2000201

2022-07-11

## Low Molecular Weight Heparin Concentration [C] in the controls

Control C3	LOT : F202900089
Target value :	0.79 IU/mL / UI/mL
Acceptance range :	[0.67-0.91] IU/mL / UI/mL

Control C4	LOT : F202800090
Target value :	1.21 IU/mL / UI/mL
Acceptance range :	[1.03-1.39] IU/mL / UI/mL

Standardization : Controls are related to International Standard 11/176 WHO from NIBSC

Approved Date / Date d'Approbation : 2020.02.06

Quality Control Manager / Responsable Contrôle Qualité : S.LECOURT

