



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : F1903098
Expiration date : 2022-04-26

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	F1903098	2022-04-26

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.5 mg/mL	≥ 80%	99%
Reactivity (% AgregMax) at 1 mg/mL	≥ 70%	87%
Reactivity (% AgregMax) at 0.5 mg/mL	< 10%	7%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-11-25

Quality Control Manager: **S. LECOURT**

