



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor IX Deficient Plasma
Product Code : DP050A/K
Lot No. : F1903091
Expiration date : 2022-04-23

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	F191803091	2022-04-23

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 80 sec	97.6 sec 57.6 sec
Concentration FIX	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	86%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2019-11-25

Quality Control Manager : **S. LECOURT**

