



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST HIA IgG,A,M

Product Code : RK040D

Lot No. : F2000099

Expiration date : 2022-04-15

Components	Quantity	Lot	Expiration Date
Heparin pre-coated plate	12x8 wells	F2000166	2022-07-22
HIA Ig G,A,M Positive Control	3 vials	F2000266	2022-08-09
Negative Control	3 vials	F1900101	2022-07-07
Anti (h) IgG - Anti IgA - Anti IgM HRP immunoconjugate	3 vials	F2000174	2023-07-31
HIA Sample diluent	2x50 ml	F1903216	2022-05-20
Wash solution	1x50 ml	F2000037	2023-06-27
Cell Lysate	3 vials	F1903067	2022-04-15
Conjugate diluent	1x25 ml	F1900898	2022-12-27
TMB	1x25 ml	190705D01	2023-07-30
Sulfuric Acid 0,45M	1x6 ml	F1903262	2023-05-29

Analytical Data	Specification	Result
Reactivity	A450 for Positive Control ≥ 1.00 A450 for Negative Control ≤ 0.30	2.29 0.14
Pathological Plasmas	A450 ≥ 0.500	PASS
Blank	< 0.100	0.032
Normal Plasma	$M + 2SD \leq 0.50$	0.27

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-03-26

Quality Control Manager: S. LECOURT

