



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL LOW
Product Code : 225101
Lot No. : F1903227
Expiration date : 2022-05-24

Components	Quantity	Lot	Expiration Date
C1	6 vials	F191603227	2022-05-24
C2	6 vials	F191603227	2022-05-24

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/mL C2 : 45 -105 ng/mL	24 ng/mL 79 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2019-12-20

Quality Control Manager: S. LECOURT

P.O.



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN™ Rivaroxaban Control Low

REF 225101

LOT F1903227



2022-05-24

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
CI LOT F191603227	ng/mL	24	14 - 34	N/A
CII LOT F191603227		79	63 - 95	