



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN™ EDOXABAN CONTROL  
**Product Code :** 225501  
**Lot No. :** F1903146  
**Expiration date :** 2022-05-10


Components	Quantity	Lot	Expiration Date
C1	6 vials	F191403146	2022-05-10
C2	6 vials	F191403146	2022-05-10

Analytical Data	Specification	Result
Concentration	C1 : 100 -200 ng/mL C2 : 250 -350 ng/mL	144 ng/mL 280 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2019-12-04

Quality Control Manager: S. LECOURT





A Company of Sysmex Group

155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
TéI : +33.1.34.40.65.10 — Fax : +33.1.34.48.72.36  
Web site : [www.hyphen-biomed.com](http://www.hyphen-biomed.com)



### BIOPHEN™ Edoxaban Control

**LOT** F1903146

**REF** 225501



2022-05-10

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>ACCEPTANCE RANGE</b>	<b>WHO STD</b>
<b>C1</b>	ng/mL	144	[ 108 - 180 ]	NA
<b>LOT F191403146</b>				
<b>C2</b>		280	[ 238 - 322 ]	
<b>LOT F191403146</b>				