



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : F1903207

Expiration date : 2022-05-24


Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F191503207	2022-05-24
CAL II	4 vials	F191503207	2022-05-24
CAL III	4 vials	F191503207	2022-05-24

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 53 101
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.00 Controls in acceptance range

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2019-12-20

Quality Control Manager: **S. LECOURT**





A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hypHEN-biomed.com



BIOPHEN™ Rivaroxaban Calibrator Low

REF 226001

LOT F1903207

 2022-05-24

WHO STD	TARGET VALUE	UNIT	
	0	ng/mL	CAL I LOT F191503207
	53		CAL II LOT F191503207
N/A	101		CAL III LOT F191503207