



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : F1903232
Expiration date : 2022-05-28

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F191303232	2022-05-28
CAL 2	4 vials	F191303232	2022-05-28
CAL 3	4 vials	F191303232	2022-05-28

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 254 497
Linearity and accuracy of calibration with values assigned	R ² ≥ 0.98 Measured values for controls in acceptance ranges	> 0.99 Measured values for controls in acceptance ranges

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-01-10

Quality Control Manager:

S. LECOURT

