



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Fibrinogen

Product Code : RK024A

Lot No. : F1903293

Expiration date : 2022-05-23

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	F2000125	2022-07-15
Fibrinogen Standard	3 vials	F1903226	2022-05-23
Fibrinogen Control I	1 vial	F1903223	2022-05-23
Fibrinogen Control II	1 vial	F1903225	2022-05-23
Fibrinogen HRP immunoconjugate	3 vials	F2000129	2022-07-18
B2F Sample diluent	2x50 ml	F2000126	2022-07-17
Wash solution	1x50 ml	F1900596	2022-10-25
B2F Conjugate diluent	1x25 ml	F2000128	2022-07-17
TMB	1x25 ml	190405D01	2023-04-30
Sulfuric Acid 0,45M	1x6 ml	F1900172	2022-07-19

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.60	1.601
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls Measured Values within acceptance ranges
Blank	< 0.150	0.016
Detection threshold	\leq 1 ng/mL	< 1 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-02-11

Quality Control Manager: S. LECOURT

