



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST FACTOR VII
Product Code : RK036A
Lot No. : F1903222
Expiration date : 2022-05-02

Components	Quantity	Lot	Expiration Date
Anti (h) Factor VII pre-coated plate	12x8 wells	F1903285	2022-06-05
Factor VII standard	3 vials	F1903131	2022-05-02
Factor VII Control I	1 vial	F1903127	2023-04-27
Factor VII Control II	1 vial	F1903128	2023-04-27
Anti-(h) Factor VII HRP immunoconjugate	3 vials	F2000005	2022-06-24
Sample diluent	2x50 ml	F1900572	2022-10-11
Wash solution	1x50 ml	F1900596	2022-10-25
Conjugate diluent	1x25 ml	F1900006	2022-06-21
TMB	1x25 ml	190405D01	2023-04-30
Sulfuric Acid 0,45M	1x6 ml	F1900172	2022-07-19

Analytical Data	Specification	Result
Reactivity	A450 for 100 % \geq 1.00	1.57
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls Measured Values within acceptance ranges
Blank	< 0.100	0.005
Detection threshold	\leq 5%	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-01-15

Quality Control Manager: S. LECOURT

