



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST Annexin V

Product Code: RK004A

Lot No.: F2000513

Expiration date: 2022-06-21

Components	Quantity	Lot	Expiration Date
Anti-(h) Annexin V pre-coated plate	12x8 wells	F2000772	2023-11-14
Annexin V standard	3 vials	F2000724	2023-11-08
Control I	1 vial	F2000726	2023-11-08
Control II	1 vial	F2000728	2023-11-08
Anti-(h)- Annexin V HRP immunoconjugate	3 vials	F2000773	2023-11-16
Sample diluent	2x50 ml	F2000031	2023-06-26
Wash solution	1x50 ml	F2000802	2023-11-20
Conjugate diluent	1x25 ml	F1900006	2022-06-21
TMB	1x25 ml	190705D01	2023-07-30
Sulfuric Acid 0,45M	1x6 ml	F2000078	2023-07-04

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.40	1.71
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls are in acceptance ranges
Blank	< 0.100	0.032
Detection threshold	\leq 1 ng/mL	< 0.1 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-06-16

Quality Control Manager: S. LECOURT