



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN a2 ANTIPLASMIN (LRT)  
**Product Code :** 220502  
**Lot No. :** F2001038  
**Expiration date :** 2022-07-24

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	F201801038	2022-07-24
R2 : SPm-41 Substrate	3 vials	F201801038	2022-07-24

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = 0.25 ± 0.10 R <sup>2</sup> ≥ 0.98 Measured values for controls in acceptance ranges	0.08 0.25 1.00 controls in acceptance ranges
Detection threshold (LOQ) on CS	≤ 10%	< 2%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2020-08-24

Quality Control Manager : **S.LECOURT**