



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description BIOPHEN ABNORMAL CONTROL PLASMA

Product Code : 223301

Lot No. : F2000200

Expiration date : 2022-07-29

Components	Quantity	Lot	Expiration Date
Abnormal Control	12 vials	F201500200	2022-07-29

Analytical Data	Specification	Result
INR	> 1.3	> 1.3
aPTT LR Ratio	> 1.50	> 1.50
Fibrinogen (Activity)	0.8-1.2 g/L	0.93 g/L
Fibrinogen (Antigen)	0.7-1.3 g/L	1.07 g/L
LA ratio	≤ 1.2	0.83
Alpha 2 Anti plasmin	20 - 50 %	33%
Antithrombin (Activity)	20 - 50 %	33%
Antithrombin (Antigen)	20 - 50 %	34%
Factor II	20 - 50 %	33%
Factor V	20 - 50 %	32%
Factor VII	20 - 50 %	31%
Factor X	20 - 50 %	32%
Factor VII+ X	20 - 50 %	32%
Factor VIII	20 - 50 %	37%
Factor IX	20 - 50 %	35%
Factor XI	20 - 50 %	27%
Factor XII	20 - 50 %	35%
Factor XIII	20 - 50 %	34%
Free Protein S (Antigen)	20 - 50 %	31%
Plasminogen	20 - 50 %	36%
Protein C Chromogenic	20 - 50 %	35%
Protein C Clotting	20 - 50 %	35%
Protein S (Activity)	20 - 50 %	35%
von Willebrand Factor (Antigen)	20 - 50 %	44%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-06-09

Quality Control Manager: **S.LECOURT**

