



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN UFH CONTROL C2  
**Product Code :** 223901  
**Lot No. :** F2000092  
**Expiration date :** 2022-07-12

Components	Quantity	Lot	Expiration Date
C2	12 vials	F201600092	2022-07-12

Analytical Data	Specification	Result
Concentration	C2 : 0.40 -0.60 IU/mL	0.48 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-02-07

Quality Control Manager: S. LECOURT