



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021A/K/L
Lot No. : F2000199
Expiration date : 2022-07-29

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	1 vial (AR021A) 3 vials (AR021K) 12 vials (AR021L)	F201700199	2022-07-29

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample ≤ 5%	0.4%
FVIII:C clotting assay	Delta CT (100%) with control sample ≤ 5%	1.6%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-02-19

Quality Control Manager : **S.LECOURT**

