



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : LA CONTROL PLASMA
Product Code : SC081K
Lot No. : F2000277
Expiration date : 2022-07-23

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201700153	2022-07-23
C2	6 vials	F201700152	2022-07-23

Analytical Data	Specification	Result
C1	Mean Normalized Ratio : 1.25 - 1.60	1.52
C2	Mean Normalized Ratio : 1.80 - 3.00	2.14

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-02-20,

Quality Control Manager : **S.LECOURT**

