



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN LRT

Product Code : 221013

Lot No. : FA027456

Expiration date : 2022-08-02

Components	Quantity	Lot	Expiration Date
R1 : SXa-11 substrate	3 vials	FA0274BH56	2022-08-02
R2 : Bovine FXa	3 vials	FA0274BI56	2022-08-02

Analytical Data	Specification	Result
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON STAR	<ul style="list-style-type: none"> For a same A405 (UFH/LMWH): MV = TV \pm0.05 IU/ml for conc. \leq 0.50 IU/ml MV = TV \pm0.10 IU/ml for conc. > 0.50 IU/ml R² \geq 0.98 	PASS
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON CS	<ul style="list-style-type: none"> For a same A405 (UFH/LMWH): MV = TV \pm0.05 IU/ml for conc. \leq 0.50 IU/ml MV = TV \pm0.10 IU/ml for conc. > 0.50 IU/ml R² \geq 0.98 	PASS
DETECTION THRESHOLD	\leq 0.05 IU/ml	< 0.01 IU/ml

Calibration Curve for information

Conc. LMWH	0	0.45	0.96	1.36	1.80
OD STAR	1.658	1.176	0.844	0.622	0.463
OD CS	0.734	0.486	0.321	0.222	0.153

	Acceptance Range	MV
C3	[0.65 - 0.89]	0.78
C4	[1.00 - 1.36]	1.20

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-02-22

Quality Control Manager : S.LECOURT

