



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : F2000237

Expiration date : 2022-08-09

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F201400237	2022-08-09
CAL 2	4 vials	F201400237	2022-08-09
CAL 3	4 vials	F201400237	2022-08-09

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/mL CAL 2 : 200 -300 ng/mL CAL 3 : 400 -600 ng/mL	0 ng/mL 260 ng/mL 527 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.99 Controls in acceptance ranges

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-03-1

Quality Control Manager: S. LECOURT

