



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	CALCIUM CHLORIDE 0.025M
Product Code :	AR001A/K
Lot No. :	F2000279
Expiration date :	2022-08-12

Components	Quantity	Lot	Expiration Date
CALCIUM CHLORIDE 0.025M	3 vials (AR001A) 12 vials (AR001K)	F201500279	2022-08-12

Analytical Data	Specification	Result
aPTT on citrated human plasma (CEPHEN)	30-40 sec	30.9
FVIII:C clotting assay	Delta CT (100%) with control sample ≤ 5%	1.0

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-02-27

Quality Control Manager : **S.LECOURT**

