



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor IX Deficient Plasma
Product Code : DP050A/K
Lot No. : F2000300
Expiration date : 2022-08-20

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	F201900300	2022-08-20

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 80 sec	93.3 sec 56.0 sec
Concentration FIX	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	84%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-03-18

Quality Control Manager: S. LECOURT

