



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN CALIBRATOR
Product Code : 222001
Lot No. : F2000361
Expiration date : 2022-09-06

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F201700361	2022-09-06
CAL 2	4 vials	F201700361	2022-09-06
CAL 3	4 vials	F201700361	2022-09-06
CAL 4	4 vials	F201700361	2022-09-06
CAL 5	4 vials	F201700361	2022-09-06

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/mL CAL 2 : 0.30 -0.50 IU/mL CAL 3 : 0.70 -0.90 IU/mL CAL 4 : 1.10 -1.30 IU/mL CAL 5 : 1.45 -1.75 IU/mL	0 IU/mL 0.40 IU/mL 0.81 IU/mL 1.20 IU/mL 1.61 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-04-06

Quality Control Manager: S. LECOURT

