



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN ORGARAN CONTROL
Product Code: 223501
Lot No.: F2000610
Expiration date: 2022-10-16

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201600610	2022-10-16
C2	6 vials	F201600610	2022-10-16

Analytical Data	Specification	Result
Concentration	C1 : 0.40 -0.60 U/mL C2 : 0.80 -1.20 U/mL	0.51 U/mL 1.02 U/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-05-15

Quality Control Manager: S. LECOURT