



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN ARIXTRA CONTROL PLASMA
Product Code: 224001
Lot No.: F2000536
Expiration date: 2022-10-08

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201400536	2022-10-08
C2	6 vials	F201400536	2022-10-08

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/mL C2: 1.05 -1.35 µg/mL	0.38 µg/mL 1.12 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-05-07

Quality Control Manager: S. LECOURT