



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN APIXABAN CONTROL  
**Product Code:** 225301  
**Lot No.:** F2000667  
**Expiration date:** 2022-10-31

| Components | Quantity | Lot        | Expiration Date |
|------------|----------|------------|-----------------|
| C1         | 6 vials  | F201900667 | 2022-10-31      |
| C2         | 6 vials  | F201900667 | 2022-10-31      |

| Analytical Data | Specification                              | Result                 |
|-----------------|--|------------------------|
| Concentration   | C1 : 140 -260 ng/mL<br>C2 : 340 -460 ng/mL | 194 ng/mL<br>387 ng/mL |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-05-28

Quality Control Manager: S. LECOURT

