


CERTIFICATE OF ANALYSIS



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

Product Description: RISTOCETIN
Product Code: AG004K
Lot No.: F2000607
Expiration date: 2022-10-14

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	F2000607	2022-10-14

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.5 mg/mL	≥ 80%	99%
Reactivity (% AgregMax) at 1 mg/mL	≥ 70%	88%
Reactivity (% AgregMax) at 0.5 mg/mL	< 10%	5%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-05-14

Quality Control Manager: S. LECOURT

