



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT™ LA-S
Product Code : CK090K
Lot No. : F2000663
Expiration date : 2022-10-30

Components	Quantity	Lot	Expiration Date
HEMOCLOT LA-S	6 vials	F201600663	2022-10-30

Analytical Data	Specification	Result
Normal Plasma	Mean CT < 45 sec NR < 1.20	28.7 0.97
Pathological Plasma	NR ≥ 1.20	1.66

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-06-09

Quality Control Manager :

S.LECOURT