



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Protein C

Product Code : RK027A

Lot No. : F2000396

Expiration date : 2022-10-28

Components	Quantity	Lot	Expiration Date
Anti (h) Protein C pre-coated plate	12x8 wells	F2000658	2022-10-28
Protein C Standard	3 vials	F2000597	2022-10-31
Protein C Control I	1 vial	F2000601	2022-10-31
Protein C Control II	1 vial	F2000598	2022-10-31
Anti-(h)-Protein C HRP immunoconjugate	3 vials	F2000721	2022-11-11
Sample diluent	2x50 ml	F2000652	2023-10-30
Wash solution	1x50 ml	F2000037	2023-06-27
Conjugate diluent	1x25 ml	F2000351	2023-09-26
TMB	1x25 ml	190705D01	2023-07-30
Sulfuric Acid 0,45M	1x6 ml	F1903262	2023-05-29

Analytical Data	Specification	Result
Reactivity	A450 for 100% \geq 1.40	1.59
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	controls are in acceptance range
Blank	< 0.100	0.02
Detection threshold	\leq 5 %	< 1%

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20	Acceptance Range	MV
OD	1.75	1.23	0.80	0.38	0.24	CI [70 - 94]	82
						CII [39 - 59]	48

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-06-09

Quality Control Manager : **S.LECOURT**