



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL LOW
Product Code : 225201
Lot No. : F2000666
Expiration date : 2022-11-01

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201800666	2022-11-01
C2	6 vials	F201800666	2022-11-01

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	24 79

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-06-04

Quality Control Manager :

S.LECOURT