



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR LOW
Product Code : 226101
Lot No. : F2000665
Expiration date : 2022-11-01

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F201700665	2022-11-01
CAL II	4 vials	F201700665	2022-11-01
CAL III	4 vials	F201700665	2022-11-01

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 53 111
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.00 controls in compliance

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-06-04

Quality Control Manager :

S.LECOURT

