



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN EDOXABAN CALIBRATOR LOW  
**Product Code :** 226401  
**Lot No. :** F2000706  
**Expiration date :** 2022-11-12

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F201300706	2022-11-12
CAL II	4 vials	F201300706	2022-11-12
CAL III	4 vials	F201300706	2022-11-12

Analytical Data	Specification	Result
Concentration	CAL I : < 15 ng/mL CAL II : 40 -60 ng/mL CAL III : 80 -120 ng/mL	0 ng/mL 50 ng/mL 112 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 > 0,99$ Measured values for controls in acceptance ranges

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-06-11

Quality Control Manager: S. LECOURT

