



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN EDOXABAN CALIBRATOR  
**Product Code :** 226501  
**Lot No. :** F2000705  
**Expiration date :** 2022-11-12

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F201400705	2022-11-12
CAL 2	4 vials	F201400705	2022-11-12
CAL 3	4 vials	F201400705	2022-11-12

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 40$ ng/mL CAL 2 : 200 -300 ng/mL CAL 3 : 400 -600 ng/mL	0 ng/mL 275 ng/mL 514 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 > 0.99$ Measured values for controls in acceptance ranges

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-06-11

Quality Control Manager: S. LECOURT

