



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** ADP  
**Product Code :** AG001K  
**Lot No. :** F2000702  
**Expiration date :** 2022-11-15

Components	Quantity	Lot	Expiration Date
ADP	3 vials	F2000702	2022-11-15

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 10 µM	70-100%	83%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2020-06-11

Quality Control Manager : **S.LECOURT**

