



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN RIVAROXABAN CALIBRATOR  
**Product Code:** 222701  
**Lot No.:** F2000903  
**Expiration date:** 2022-12-17

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F201500903	2022-12-17
CAL 2	4 vials	F201500903	2022-12-17
CAL 3	4 vials	F201500903	2022-12-17

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 50$ ng/mL CAL 2 : 200 -300 ng/mL CAL 3 : 400 -600 ng/mL	0 ng/mL 261 ng/mL 503 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$> 0.98$ Controls are in acceptance ranges

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-07-20

Quality Control Manager:

**S. LECOURT**

