



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN RIVAROXABAN CONTROL LOW
Product Code: 225101
Lot No.: F2000902
Expiration date: 2022-12-17

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201700902	2022-12-17
C2	6 vials	F201700902	2022-12-17

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/mL C2 : 45 -105 ng/mL	26 ng/mL 81 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-07-20

Quality Control Manager: S. LECOURT