



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST Fibrinogen
Product Code: RK024A
Lot No.: F2000739
Expiration date: 2022-12-11

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	F2000940	2022-12-16
Fibrinogen Standard	3 vials	F2000884	2022-12-11
Fibrinogen Control I	1 vial	F2000887	2022-12-11
Fibrinogen Control II	1 vial	F2000888	2022-12-11
Fibrinogen HRP immunoconjugate	3 vials	F2000950	2022-12-23
B2F Sample diluent	2x50 ml	F2000948	2022-12-23
Wash solution	1x50 ml	F2000802	2023-11-20
B2F Conjugate diluent	1x25 ml	F2000949	2022-12-23
TMB	1x25 ml	190705D01	2023-07-30
Sulfuric Acid 0,45M	1x6 ml	F2000078	2023-07-04

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.60	1.60
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Control in acceptance range
Blank	< 0.150	0.016
Detection threshold	\leq 1 ng/ml	< 0.1

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-07-20

Quality Control Manager: S. LECOURT

