



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN LMWH CONTROL LOW  
**Product Code:** 223701  
**Lot No.:** F2001196  
**Expiration date:** 2023-01-14

Components	Quantity	Lot	Expiration Date
CI	6 vials	F201A01022	2023-01-14
CII	6 vials	F201801023	2023-01-14

Analytical Data	Specification	Result
Concentration	CI: 0.15 -0.35 IU/mL CII: 0.35 -0.65 IU/mL	0.27 0.52

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-08-31

Quality Control Manager: S. LECOURT