



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH CONTROL LOW CI
Product Code: 224301
Lot No.: F2001022
Expiration date: 2023-01-14

Components	Quantity	Lot	Expiration Date
CI	12 vials	F201A01022	2023-01-14

Analytical Data	Specification	Result
Concentration	CI: 0.15 -0.35 IU/mL	0.27

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-08-31

Quality Control Manager: S. LECOURT