



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH CONTROL LOW CII
Product Code: 224401
Lot No.: F2001023
Expiration date: 2023-01-14

Components	Quantity	Lot	Expiration Date
CII	12 vials	F201801023	2023-01-14

Analytical Data	Specification	Result
Concentration	CII: 0.35 -0.65 IU/mL	0.52

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-08-31

Quality Control Manager: S. LECOURT