



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN FVIIa CONTROL SET
Product Code : 224901
Lot No. : F2001026
Expiration date : 2023-01-16

Components	Quantity	Lot	Expiration Date
C1	3 vials	F201401026	2023-01-16
C2	3 vials	F201401026	2023-01-16

Analytical Data	Specification	Result
Concentration	C1 : 50 -100 mIU/ml C2 : 200 -300 mIU/ml	86 mIU/mL 275 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-08-24

Quality Control Manager : S.LECOURT