



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor VIII:C Deficient Plasma  
**Product Code :** DP040A/K  
**Lot No. :** F2001027  
**Expiration date :** 2023-01-16

| Components                     | Quantity                            | Lot        | Expiration Date |
|--------------------------------|-------------------------------------|------------|-----------------|
| Factor VIII:C Deficient Plasma | 1 vial (DP040A)<br>6 vials (DP040K) | F201901027 | 2023-01-16      |

| Analytical Data     | Specification                                   | Result                             |
|---------------------|---|------------------------------------|
| Calibration curve   | CT (0-100%) $\geq$ 40 sec<br>CT (100%) < 75 sec | 53.7 sec (STAR)<br>56.8 sec (STAR) |
| Concentration FVIII | < 1%  | < 1%                               |
| Concentration FV    | > 50 %  | 83%                                |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2020-08-24

Quality Control Manager : **S.LECOURT**