




Product name:	Protein L ligand leakage ELISA kit		
Production date:	2022-02		
Date of release:	2022-06-17		
Expiry date:	2023-02		
Store:	+4°C. Protein L reference at -20°C or -80°C.		
Contents:	ELISA kit for detection and quantification of ligand leakage in eluates from Protein L affinity chromatography medium containing immunoglobulin (Ig) or Ig-fragments.		
	Kit contents	Lot	Expiry date
	1 x Anti-Protein L coated microtiter plate	280615	2023-02
	130 µl HRP-conjugated anti-Protein L pAb	282004	2023-05
	200 µl Protein L reference (600 ng/ml)	281409-01	2024-04
	13 ml EC-Blue Enhanced™	282007	2023-09
	7 ml Stop solution, 0.5 M H ₂ SO ₄	271623	2026-04
	6 x 5 g PBS Tween tablets	281701-02	2025-04

Quality Control

Analysis	Specification	Result
%CV	< 12 %	Passed
Absorbance 30 ng/ml	> 2.1	Passed

	OD 450 nm	Concentration
Calibrator 0.47 ng/ml	0,188	-
Calibrator 0.94 ng/ml	0,233	-
Calibrator 1.88 ng/ml	0,465	-
Calibrator 3.75 ng/ml	0,806	-
Calibrator 7.5 ng/ml	1,526	-
Calibrator 15 ng/ml	2,331	-
Calibrator 30 ng/ml	2,606	-

Control sample A	0,227	0,84 ng/ml
Control sample B	0,800	3,63 ng/ml
Control sample C	1,750	8,96 ng/ml



The above material has met all quality specifications and has been reviewed by a quality representative

Max Johansson
 Quality Assurance, Max Johansson

2022-06-20
 Date