



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH CONTROL PLASMA
Product Code: 223001
Lot No.: F2001195
Expiration date: 2023-02-04

Components	Quantity	Lot	Expiration Date
C3	6 vials	F202A01079	2023-02-04
C4	6 vials	F202901080	2023-02-04

Analytical Data	Specification	Result
Concentration	C3: 0.70 -0.90 IU/mL C4: 1.05 -1.35 IU/mL	0.77 IU/mL 1.19 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-08-31

Quality Control Manager: S. LECOURT