



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN LMWH CONTROLC3  
**Product Code:** 223801  
**Lot No.:** F2001079  
**Expiration date:** 2023-02-04

Components	Quantity	Lot	Expiration Date
C3	12 vials	F202A01079	2023-02-04

Analytical Data	Specification	Result
Concentration	C3: 0.70 -0.90 IU/mL	0.77 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2020-08-31

Quality Control Manager: S. LECOURT