



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: Factor VII Deficient Plasma
Product Code: DP030A/K
Lot No.: F2001148
Expiration date: 2023-02-21

Components	Quantity	Lot	Expiration Date
Factor VII Deficient Plasma	1 vial (DP030A) 6 vials (DP030K)	F201501148	2023-02-21

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 100 sec CT (100%) < 40 sec Thromboplastin : NEOPTIMAL STAGO	110 28
Concentration FVII	< 1%	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-09-10

Quality Control Manager: S. LECOURT