



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST PF4
Product Code : RK006A
Lot No. : F2001151
Expiration date : 2023-02-17

Components	Quantity	Lot	Expiration Date
PF4 pre-coated plate	12x8 wells	F2001170	2023-02-24
PF4 Standard	3 vials	F2001166	2023-02-17
PF4 Control I	1 vial	F2001167	2023-02-17
PF4 Control II	1 vial	F2001169	2023-02-17
PF4 HRP immunoconjugate	3 vials	F2001174	2023-03-05
PF4 Sample diluent	2x50 ml	F2001172	2023-02-26
Wash solution	1x50 ml	F2000802	2023-11-20
Conjugate diluent	1x25 ml	F2000352	2023-09-26
TMB	1x25 ml	200204D01	2024-02-28
Sulfuric Acid 0,45M	1x6 ml	F2000078	2023-07-04

Analytical Data	Specification	Result
Reactivity	A450 for 8 IU/ml \geq 1.70	1.86
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls are in acceptance range
Blank	< 0.100	0.065
Detection threshold	\leq 0.5 IU/ml	< 0.1 UI/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2020-10-05

Quality Control Manager :

S.LECOURT
