



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST TOTAL TF

**Product Code :** RK042A

**Lot No. :** FA0185

**Expiration date :** 2023-02-25

Components	Quantity	Lot	Expiration Date
Anti-(h)-TF pre-coated plate	12x8 wells	F2001156	2023-02-25
TF Standard	3 vials	F2001157	2023-03-13
Biotinylated Antibody	1 vial	F2001191	2023-03-20
Control I	1 vial	F2001159	2023-03-13
Control II	1 vial	F2001160	2023-03-13
HRP streptavidine immunoconjugate	3 vials	F2001192	2023-03-27
TF-Sample diluent	2x40 ml	FA0184	2023-07-01
Wash solution	1x50 ml	F2000802	2023-11-20
TF - Conjugate diluent	1x25 ml	F2001162	2023-03-05
TMB HS	1x25 ml	200517D04	2024-05-31
Sulfuric Acid 0,45M	1x6 ml	F2000078	2023-07-04

Analytical Data	Specification	Result
Reactivity	A450 for 500 pg/ml $\geq$ 1.60	1.74
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls Measured Values within acceptance ranges
Blank	< 0.150	0.089
Detection threshold	$\leq$ 10 pg/ml	< 2

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-01-25

Quality Control Manager : S.LECOURT