



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN a2 ANTIPLASMIN (LRT)
Product Code : 220502
Lot No. : FA06171A
Expiration date : 2023-03-12

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	FA0617171A	2023-03-12
R2 : SPm-41 Substrate	3 vials	FA0617181A	2023-03-12

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05	0.09
	A405/min(100%) = 0.25 ± 0.10	0.26
	R ² ≥ 0.98	1.00
	Measured values for controls in acceptance ranges	Controls are in acceptance ranges
Detection threshold (LOQ) on CS	≤ 10%	< 0.1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-04-09

Quality Control Manager : S.LECOURT