



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN DABIGATRAN CONTROL PLASMA LOW
Product Code: 225001
Lot No.: F2001291
Expiration date: 2023-03-21

Components	Quantity	Lot	Expiration Date
C1	6 vials	F201501291	2023-03-21
C2	6 vials	F201501291	2023-03-21

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/mL C2 : 45 -105 ng/mL	23 ng/mL 67 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-10-15

Quality Control Manager: S. LECOURT

S. Lecourt