



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT FACTEUR VIIa
Product Code : CK092K
Lot No. : F2000483-F2000487
Expiration date : 2023-03-22

Components	Quantity	Lot	Expiration Date
R1: Factor VII Deficient Plasma	3 vials	F201900483 / F201200487	2023-03-23
R2 : rTTF and PLPs	3 vials	F201900483 / F201200487	2023-04-01
R3 : Hepes BSA Buffer	3 vials	F201900483 / F201200487	2023-03-22

Analytical Data	Specification	Result
Calibration curve on CS	Delta CT (500-12.5 mIU/mL) \geq 60 sec	78.5 sec
	CT (250 mIU/mL) = 40 ± 7 sec	36.8 sec
	$R^2 \geq 0.98$	0.999
	Normal Plasma : 15-140 mIU/mL	39.0 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2020-05-14

Quality Control Manager: S. LECOURT