



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ BIVALIRUDIN CONTROL
Product Code : 225701
Lot No. : F1801436 / F1801435
Expiration date : 2021-04-13

Components	Quantity	Lot	Expiration Date
C1	3 vials	F181101436 F181101435	2021-04-13
C2	3 vials	F181101436 F181101435	2021-04-13

Analytical Data	Specification	Result
Concentration	C1 : 1.25 -1.75 µg/ml C2 : 3.3 - 4.3 µg/ml	1.37 µg/mL 3.62 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-05-10

Quality Control Manager :

C.DUNOIS





BIOPHEN™ Bivalirudin Control

REF 225701-RUO

LOT F1801435

2021-04-13

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C1 LOT F181101435	µg/mL	1.37	[1.12 – 1.62]	NA
C2 LOT F181101435		3.62	[2.90 – 4.34]	